



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

56

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
09/599,674	06/22/2000	Brian J. Brown	S63.2-9216	7508

490 7590 02/25/2004

VIDAS, ARRETT & STEINKRAUS, P.A.
6109 BLUE CIRCLE DRIVE
SUITE 2000
MINNETONKA, MN 55343-9185

EXAMINER

PREBILIC, PAUL B

ART UNIT	PAPER NUMBER
----------	--------------

3738

DATE MAILED: 02/25/2004

17

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary

Application No.

09/599,674

Applicant(s)

BROWN ET AL.

Examiner

Paul B. Prebilic

Art Unit

3738

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 10 December 2003.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 9-14 and 17-20 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 9-14 and 17-20 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
 - ☐ Certified copies of the priority documents have been received in Application No. _____.
 - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- ☐ Notice of References Cited (PTO-892)
- ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- ☒ Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
Paper No(s)/Mail Date 15.
- ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____.
- ☐ Notice of Informal Patent Application (PTO-152)
- ☐ Other: _____.

Information Disclosure Statement

The information disclosure statement filed November 12, 2003 fails to comply with 37 CFR 1.98(a)(2), which requires a legible copy of each U.S. and foreign patent; each publication or that portion which caused it to be listed; and all other information or that portion which caused it to be listed. The copies of the two patent applications were not provided so they could not be considered. Therefore, they were struck from the PTO-1449 form. Applicant is respectfully requested to provide copies of such in response to this office action.

Claim Rejections - 35 USC § 112

The following is a quotation of the first paragraph of 35 U.S.C. 112:

The specification shall contain a written description of the invention, and of the manner and process of making and using it, in such full, clear, concise, and exact terms as to enable any person skilled in the art to which it pertains, or with which it is most nearly connected, to make and use the same and shall set forth the best mode contemplated by the inventor of carrying out his invention.

Claims 9-14 and 17 are rejected under 35 U.S.C. 112, first paragraph, as failing to comply with the written description requirement. The claim(s) contains subject matter, which was not described in the specification in such a way as to reasonably convey to one skilled in the relevant art that the inventor(s), at the time the application was filed, had possession of the claimed invention. The new language of "the first type of set of strut members having a shorter perimeter than a closed member formed by a second type of set of strut members" lacks original support. This is due to the fact that a perimeter is defined as "A closed curve bounding a plane area" American Heritage Dictionary, Second College Edition, (1982). Since a perimeter is supposed to be in one plane, then the perimeter of each cylindrical section of the disclosed stent is the same in

Art Unit: 3738

all embodiments disclosed; see the figures. For this reason, this description is not supported by the original disclosure, which only discloses all perimeters as being the same for all cylindrical sections.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(e) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

Claims 9-14 and 17-20 are rejected under 35 U.S.C. 102(e) as being anticipated by Simon et al (US 5,354,308) where the first type of strut members as claimed are the zigzag ring of struts shown along the radius of where reference number 14 points in Simon, and the second type of struts as claimed are the ring of struts with fingers (16) at the ends of the stent of Simon; see Figures 4 and 5 as well as column 3, line 9 to column 4, line 27. In Figure 5, the struts are parallel to each other.

With regard to claims 12 and 14, in the cooled soft state, the Nitinol material of Simon is ductile, and thus, balloon expandable to the extent required by the present claims.

With regard to claims 17 and 19, the middle section of Simon (see Figures 3 or 4) can be said to contain the connectors, and some of the struts of the middle section

Art Unit: 3738

are oblique with respect to the longitudinal axis. For this reason, the claim language is fully met.

Response to Arguments

Applicant's arguments filed November 12, 2003 have been fully considered but they are not persuasive.

In response to the traversal that Simon discloses a non-uniform diameter when expanded citing Figure 1 and column 3, lines 9-15, the Examiner asserts that Figure 1 shows a uniform diameter stent and column 3, lines 9-15 state nothing about a non-uniform diameter. Rather, Figures 3, 4, 7C, and 8C show stents with uniform diameters. For this reason, the traversal is considered unpersuasive.

Conclusion

THIS ACTION IS MADE FINAL. Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire **THREE MONTHS** from the mailing date of this action. In the event a first reply is filed within **TWO MONTHS** of the mailing date of this final action and the advisory action is not mailed until after the end of the **THREE-MONTH** shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than **SIX MONTHS** from the mailing date of this final action.

Applicant should specifically point out the support for any amendments made to the disclosure, including the claims (MPEP 714.02 and 2163.06). Due to the procedure

Art Unit: 3738

outlined in MPEP 2163.06 for interpreting claims, it is noted that other art may be applicable under 35 USC 102 of 35 USC 103(a) once the aforementioned issue(s) is/are addressed.

Applicant is respectfully requested to provide a list of all copending applications that set forth similar subject matter to the present claims. A copy of such copending claims is respectfully requested in response to this Office action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Paul Prebilic whose telephone number is (703) 308-2905. The examiner can normally be reached on Monday-Thursday from 6:30 AM to 5:00 PM.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Corrine McDermott, can be reached on (703) 308-2111. The fax phone number for this Technology Center is (703) 872-9306.

Any inquiry of a general nature or relating to the status of this application should be directed to the Technology Center 3700 receptionist whose telephone number is (703) 308-0858.



Paul Prebilic
Primary Examiner
Art Unit 3738